



9620 Medical Center Drive
Rockville, MD 20850, USA
www.abpbiotech.com.cn

CERTIFICATE OF ANALYSIS

Catalog Number N011
Product Name iQuant™ dsDNA HS Assay Kit
Lot Number AB1710N1

	LOT DATA	SPECIFICATION
Absorption Optical Density of Component D	0.204 at 260 nm	0.2 ± 0.008 at 260 nm
Limit of Detection Sensitivity	0.1 ng/well	≤ 0.2 ng/well
Linear Regression R-squared	0.999	≥ 0.98